

**URGENT MEDICAL DEVICE CORRECTION**  
**All CADD-Solis™ 21-2160-XX\* Li-ion Rechargeable Battery Packs**

16 July 2024:

Dear Valued CADD-Solis Customers:

- Director of Biomedical Engineering
- Director of Nursing
- Director of Risk Management

Smiths Medical is issuing this letter to notify you of an issue with all CADD-Solis Li-ion Rechargeable Battery Packs (List Number 21-2160-XX\*). The lithium-ion battery packs provide an alternate source of power for the CADD-Solis ambulatory infusion pump.

Smiths Medical is notifying all affected CADD-Solis customers of this issue for awareness. This notification details the issue and the affected product models.

**Affected Models:**

**Table 1: Affected Products(s)**

Product Name	List Number
<b>CADD-Solis Li-ion Rechargeable Battery Packs</b>	21-2160-XX*

\*Please note that the XX suffix is region specific.

**Overview of the Issue:**

Smiths Medical identified three (3) reports in which damage to the battery pack may have caused a short to a capacitor within the battery pack. While the battery encasement is designed to be flame retardant, a short to the capacitor may potentially lead to melting of the battery pack case. If this issue occurs, the battery pack charging circuit may become inoperable.

**Potential Risk:**

Damage to the battery pack may lead to a delay in therapy or interruption of therapy. The user would be alerted with the normal “Low Battery” or “Depleted Battery” alarms. The presence of excessive heat in the event of a melted battery pack casing is also possible, which may result in a thermal injury.

**To date, Smiths Medical has not received any reports of serious injuries or deaths related to this issue.**

**Actions for Users:**

Inform all affected CADD Solis users of rechargeable battery packs of this notice. Provide the instructions below:

1. Examine the external condition of the battery pack and look for evidence of damage to the outer case. As stated in the battery pack Instructions for Use, if the battery pack housing is cracked or otherwise damaged, replace the battery pack. NEVER use a battery pack that appears damaged. A rechargeable battery pack must be replaced with either another CADD-Solis rechargeable battery pack or with 4 AA batteries.
2. Users with damaged battery packs should submit complaints per the contact information below.
3. Ensure all users or potential users of these products are immediately made aware of this notification.

4. Complete and return the attached Response Form to [smithsmedical5437@sedgwick.com](mailto:smithsmedical5437@sedgwick.com) **within ten days of receipt** to acknowledge your understanding of this notification, even if you do not have the affected product. Please contact your local representative for credit.
5. **DISTRIBUTORS:** If you have distributed potentially affected products to your customers, please immediately forward this notice to them. Request that they complete the response form and return it to [smithsmedical5437@sedgwick.com](mailto:smithsmedical5437@sedgwick.com)

**Follow up Actions by Smiths Medical:**

Smiths Medical is continuing to investigate this matter to determine if additional actions may be warranted.

For further inquiries, please contact Smiths Medical using the following information:

Smiths Medical Contact	Contact Information	Areas of Support
Global Complaint Management	<a href="mailto:globalcomplaints@icumed.com">globalcomplaints@icumed.com</a> 1-866-216-8806	To report adverse events or product complaints
Technical Support	<a href="mailto:TSC.Support@icumed.com">TSC.Support@icumed.com</a> 1-800-241-4002, option 3	Additional information or technical assistance

This notification is being performed with the knowledge of regulatory authorities, including the US Food and Drug Administration (FDA).

If you wish to contact the FDA regarding any adverse events or quality problems associated with this notice, use the following contact information.

- [www.fda.gov/medwatch](http://www.fda.gov/medwatch)
- 1-(888)-INFO-FDA

Smiths Medical is committed to patient safety and is focused on providing exceptional product reliability and the highest level of customer satisfaction. Thank you for your prompt support on this important matter. We appreciate your cooperation.

Sincerely,



Jim Vogel  
Vice President of Quality

**Enclosures:**

- Response Form (separate document)